

## LiQuilaz® S05 Thermal Test

### Introduction

Optical particle counters like the LiQuilaz® S05 are used in a variety of environments which may be subject to changes in ambient temperatures. In addition, the LiQuilaz itself generates heat due to warming of its internal components. Particle Measuring Systems (PMS) tested the S05 in an environmental chamber to determine if true particle counts are influenced by temperature when measuring ambient distribution of particles in water. Tests with calibration particles were also conducted in order determine the effect of temperature on sizing accuracy.

### Environmental Test

Two S05 sensors were connected in parallel to an ultrapure water (UPW) source. A variable speed particle injector (VSPI) was connected upstream of the sensors in order to introduce a fixed concentration of particle contamination. The sensor's case temperature was measured using a thermocouple taped to the top of the lid on the outside of both sensors. One S05 sensor was mounted inside an environmental chamber and the second sensor remained outside the chamber to serve as a control. The environmental chamber reaches thermal equilibrium in about one minute. The chamber's temperature was increased from 22°C to 40°C by raising the temperature setting every 15 minutes in increments of approximately 3°C. The temperature of the control S05 remained relatively constant during the test. DC light stayed at 0.010 volts throughout the test at PMS on both sensors. DC light measurements correlate to increase in background scattering.

Particle concentrations measured by the control S05 remained constant throughout the test at 57 cts/mL +/- 1.2 cts/mL Particle concentrations measured by the S05 in the chamber remained constant at 54 cts/mL +/- 0.8 cts/mL. The results show excellent agreement as documented in Table 1. Absolute differences between the sensors are insignificant and well below the expected matching range of 20%. The plots of both sensors are shown in Figure 1.

<b>Table 1. LiQuilaz S05 temperature test at Particle Measuring Systems</b>				
Environmental Chamber Temperature (°C)	S05 Sensor Under Test		S05 Control Sensor	
	Case Temperature (°C)	Particle Concentration (cts/mL)	Case Temperature (°C)	Particle Concentration (cts/mL)
22	23.4	54.8	28.4	58.2
25	25.5	52.8	28.0	58.0
28	28.2	53.4	27.7	56.8
31	31.3	54.1	27.4	57.0
35	35.1	53.7	27.4	55.8
40	40.0	55.0	27.3	55.0
<b>Average</b>	<b>54.0</b>		<b>56.8</b>	
<b>Std. Dev.</b>	<b>0.8</b>		<b>1.2</b>	





Figure 1. Particle concentrations for LiQuilaz S05 under test and S05 control at various environmental chamber temperatures. The plot demonstrates the independence of true particle counts with change in temperature.

## Internal Sensor Temperature

The LiQuilaz sensor takes approximately 90 minutes to reach a stable internal temperature of 30°C, as shown in Figure 2, when operated in an ambient temperature of 23°C. The internal temperature was measured with a thermocouple located on top of the S05 power supply. The internal temperature increase in Figure 2 occurs due to the heat being generated by the S05's electronics.

Although it does not affect data interpretation, internal temperatures may still have been increasing slightly in the previous environmental test. Longer stabilization times (90 minutes vs. 30 minutes) may have caused an additional 5-10% increase in internal



temperatures. In order to demonstrate operation throughout the temperature range of the sensor, the temperature was raised to 40°C in the above test (5°C above specification) with no affect on particle counts. Further results from preliminary testing using internal and external sensors are discussed below.




Figure 2. Thermal equilibrium occurs after 90 minutes when operated in an ambient environment of 23°C

### **Preliminary testing monitoring internal temperature**

A preliminary S05 temperature test was performed in the environmental chamber using a similar setup as described above. Temperatures were increased from 22°C to 35°C and back down to 15°C. Particle concentrations decreased with time throughout the test and showed no relationship to temperature. Temperatures were measured inside and outside the S05 sensor as described above. Temperatures inside the LiQuilaz were recorded after about 30 minutes with a typical delta of about 3.5°C. This data is shown in Figure 3.

The particle concentration data shows the SO5 is insensitive to temperature changes. The downward drift in the data is a result of the VSPI and plumbing cleaning up. In addition, zero count measurements were successfully made at 15°C and 35°C on UPW, indicating that electronic noise does not influence particle counts over the temperature range.

**S05 Thermal Test  
July 20, 2006**




Figure 3. Particle concentrations at various casing temperatures. This demonstrates the independence of true particle counts and temperature. The decreasing counts are caused by the plumbing becoming clearer.

### Particle Sizing

Measuring the sizing errors is one of the best ways to verify that the sensor is working properly. Particle Measuring Systems has developed a spreadsheet tool that calculates sizing error based upon the measurement of calibration polystyrene latex spheres (PSL). Sizing errors are typically less than 5% for a sensor with good calibration. A calculation of sizing errors in elevated temperatures is presented in Table 2. Temperature has a negligible impact on sizing error.

Table 2. 0.5 micron PSL Testing at temperature			
	Counts Measured	Split	Sizing error
Control S05 Room Temp	337	43.5%	-0.82%
Control S05 Room Temp	315	40.6%	-1.19%
Control S05 Room Temp	326	42.1%	-1.00%
S05 Test Room Temp	331	42.7%	-0.92%
S05 Test 35°C	379	48.9%	-0.14%
S05 Test 30°C	378	48.8%	-0.14%
S05 Test Room Temp	357	46.1%	-0.49%
Total counts	775		

## Conclusion

True particle counts in water for ambient distributions showed no effect when changing temperature over the normal operating range of the S05 sensor. Measurements of calibration particles showed no significant difference in sizing error at elevated temperatures. The measured sizing errors were 1% or less. This is significantly less than typical sizing errors of 3 -5% for a calibrated sensor. The LiQuilaz S05 is a robust sensor capable of making very repeatable particle measurements over a wide temperature range.

Ap 73

Prepared by Ed Terrell, Roger Carlone, Jerry Gromala

LiQuilaz ® is a registered trademark of Particle Measuring Systems, Inc.

© Copyright Particle Measuring Systems. All rights reserved.

Reproduction or translation of any part of this work without the permission of the copyright owner is unlawful. Requests for permission or further information should be addressed to Particle Measuring Systems, Inc. at 1-800-238-1801



Particle Measuring Systems Headquarters  
5475 Airport Blvd., Boulder, CO 80301  
1-303-443-7100 1-800-238-1801 Fax: 1-303-546-7331  
Instrument Service & Support: 1-800-557-6363  
Customer Response Center: 1-877-475-3317

Particle Measuring Systems Europe  
Tel: +44 1684 581000  
PMSEurope@pmeasuring.com

Particle Measuring Systems Japan  
PMSJapan@pmeasuring.com

Particle Measuring Systems Asia Pacific  
PMSAsiaPacific@pmeasuring.com

Particle Measuring Systems Singapore  
Tel: +65-6496 0330  
PMSSingapore@pmeasuring.com

Particle Measuring Systems China  
PMSChina@pmeasuring.com

Particle Measuring Systems Mexico  
PMSMexico@pmeasuring.com

Particle Measuring Systems Puerto Rico  
PMSPuertoRico@pmeasuring.com

© 2008 Particle Measuring Systems, Inc. All rights reserved.