

	USP<797> Requirements Handilaz Mini Operations Documents	Revision date: 19 December 2006
		Revision: 1.0
		Page 1 of 38

Table of Contents

Overview	2
Sample Pharmacy USP<797> Operating Documents	3
1. Pharmacy - Certification of IV Preparation Hoods, Isolators and Cleanroom HEPA Filters	4
2. Pharmacy - Personnel Training and Evaluation of Aseptic Technique	5
3. Pharmacy - Use of Laminar Flow Hoods and Isolators	6
4. Pharmacy – Sterile Compounding Quality Assurance Program	7
5. Pharmacy – Buffer and Cleanroom Procedure	8
6. Pharmacy – Aseptic Technique	9
7. Pharmacy – Cleaning and Sanitization of the Buffer and Cleanroom Area	10
8. Pharmacy – Environmental Monitoring Program	12
9. Pharmacy - Personnel Assessment Checklist	14
Equipment Operating Procedure for the Handilaz Mini	18
1. Starting the Particle Counter	19
2. Sampling	19
3. Changing Sampling Settings	19
4. Measure Mode Settings	20
5. Using REPEAT Sampling Mode	20
6. SINGLE Sampling Mode	21
7. Using CONT (Continuous) Sampling Mode	22
8. CALC (Calculation) Mode	23
9. ISO Sampling Mode	23
10. DISPLAY Data	24
11. UPLINK Data	25
12. PRINT Data	25
13. Setting ALARMS	26
Menu Tree	27
Abbreviations and Definitions	28
Sample Standard Operating Procedure for the Handilaz Mini	31
Cleanroom Certification to ISO Class 5 & 7 Worksheets	36
Certifying a Cleanroom to ISO 14644-1, Class 5	37
Certifying a Cleanroom to ISO 14644-1, Class 7	38

 PARTICLE MEASURING SYSTEMS	USP<797> Requirements Handilaz Mini Operations Documents	Revision date: 19 December 2006
		Revision: 1.0
		Page 2 of 38

Overview

This document provides sample documents describing high-level pharmacy requirements for compliance with the regulation *USP <797> Pharmaceutical Compounding – Sterile Preparations*. Additionally, this document provides an Equipment Operating Procedure for using the Particle Measuring Systems' *Handilaz Mini*, handheld aerosol particle counter, to perform routine monitoring of cleanroom areas. Finally, worksheets for calculating ISO14644-1 Class 5 & 7 cleanroom certification requirements are provided.