

CERTIFICATE OF ANALYSIS Nr 575



Product:	BioCapt® Single-Use Manufactured by CPC Biotech srl			Product Code: 79001209			
Media:		E SOYA AGAR + NEUTR Soya Agar + 0.07% Lecith			05% Sodic	othiosulfate + 0.19	% Histidine)
Lot Nr: 18H0703		Date of Manufacturing 07 Aug 2018	J:	Expiry date : 07 Aug 2019		Storage Condit 2-25°C up to ex	
Sterilization: ≥25 KGy gamma irradiated Irradiation Certificate Nr: 6461				L	Conforms		
CHEMICAL /	PHYSICAL	TESTS:					
Appearance: light amber coloured, clear to slightly opalescent gel forms in BioCapt device plates					Conforms		
pH specification: 7.3±0.2 at 25°C pH result: 7.1				Conforms			
Filling media Volume: 30 ml ± 1ml					Conforms		
Media Comp	osition: TSA	+ neutralizers – BSE fi	r ee (s	see tecnichal data sl	neet cod 79	9001209)	Conforms

MICROBIOLOGICAL TEST CONTROLS:

Sterility Control - Absence of contamination	
At the following incubation conditions (performed before gamma irradiation): 32.5 ± 2.5°C aerobic for 7 days	Conforms
22.5 ± 2.5°C aerobic for 7 days	

Fertility - Growth Promotion test (performed after gamma irradiation):					
Control strain	Medium inoculation level	Incubation Conditions	Recovery Specifications	Recovery Results	Cultural Response
<i>E. coli</i> ATCC 8739	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	90.2%	Conforms
<i>P. aeruginosa</i> ATCC 9027	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	96.2%	Conforms
<i>S. aureus</i> ATCC 6538	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	96.5%	Conforms
<i>B. subtilis</i> ATCC 6633	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	93.3%	Conforms
<i>C. albicans</i> ATCC 10231	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	93.3%	Conforms
<i>A. brasiliensis</i> ATCC 16404	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	100.0%	Conforms

GROWTH PROMOTION TEST IN ACCORDANCE TO HARMONISED EP – USP PHARMACOPOEIAS CHAPTER, BY QUANTITATIVE SURFACE PLATING METHOD. **RECOVERY:** RESULTS AS CFU ENUMERATION ON TEST BATCH AGAR PLATES X 100/CFU ENUMERATION ON REFERENCE TESTS.

REFERENCE MEDIUM FOR RECOVERY TESTS: TRYPTIC SOYA AGAR FOR ALL BACTERIA AND SABOURAUD DEXTROSE AGAR FOR YEAST & MOLD SPECIES. **ATCC** IS REGISTERED TRADE MARK OF AMERICAN TYPE CULTURE COLLECTION.

Lot status	Approved	Date 08 Oct 2018		
Quality Control	Stefania Nichele	Nichele Stefanio		
Quality Assurance	Celeste Annovazzi	Geste Suprasi		
CPC Biotech s.r.l. Via per Ossona, 22	tel. +39 039 6898029	PMS 79001209		
20010 Casorezzo (MI) Italia www.cpcbiotech.it	fax. +39 039 6899774	Rev. 01 _18/06/2018		