

**Certificate of Analysis** 



Lot status: APPROVED

MODEL:	BioCapt single use manufactured by CPC Biotech srl	PRODUCT CODE:	79001209
MEDIA:	Tryptone Soya Agar + 0.07% Lecithin + 0.5% Tween80 + 0.1% L-Hystidine + 0.05% S-Thiosulphate		
BATCH:	18L1906	STORAGE CONDITIONS:	at 2-25°C until expiry date
DATE OF MA	ANUFACTURING: 19 Dec 2018	EXPIRY DATE:	19 Dec 2019

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	light amber coloured, clear to slightly opalescent gel forms in BioCapt device plates	Conforms
pH specification	7.3±0.2	Conforms
Filling media volume	30ml±1ml	Conforms
Media composition	TSA+NEUTRALIZERS (Medium is prepared utilizing raw materials declared to be BSE FREE by the manufacturer)	Conforms
Packaging	Triple Wrapped Sterile Irradiated (T.W.S.I.)	Conforms
Dose of irradiation	≥25 KGy	Conforms Irradiation certificate Nr: 445

## MICROBIOLOGICAL TEST CONTROLS

## • <u>Sterility control</u> (performed before gamma irradiation)

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 168 hours 22.5±2.5°C aerobic for 168 hours	No growth	Conforms

## • <u>Fertility – Growth Promotion Test</u> (performed after gamma irradiation)

Control strain	Medium inoculation level	Incubation Conditions	Recovery Specifications	Recovery Results	Cultural Response
E. coli ATCC 8739	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	81.7%	Conforms
P. aeruginosa ATCC 9027	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	73.8%	Conforms
<i>S. aureus</i> ATCC 6538	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	90.2%	Conforms
<i>B. subtilis</i> ATCC 6633	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	96.7%	Conforms
<i>C. albicans</i> ATCC 10231	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	89.1%	Conforms
<i>A. brasiliensis</i> ATCC 16404	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	89.8%	Conforms
Environmental strain #1	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	88.7%	Conforms
Environmental strain #2	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	88.9%	Conforms

GROWTH PROMOTION TEST IN ACCORDANCE TO HARMONISED EP – USP PHARMACOPOEIAS CHAPTER, BY QUANTITATIVE SURFACE PLATING METHOD.

**Recovery:** Results as CFU enumeration on test batch agar plates x 100/CFU enumeration on reference tests.

**REFERENCE MEDIUM** FOR RECOVERY TESTS: TRYPTONE SOYA AGAR FOR ALL BACTERIA AND SABOURAUD DEXTROSE AGAR FOR YEAST & MOLD SPECIES. **ATCC** IS REGISTERED TRADE MARK OF AMERICAN TYPE CULTURE COLLECTION.

Date analisys start	19 Dec 2018	Date analisys end	01 Feb 2019
Quality Control	Stefania Nichele	Nichele Stefanio	
Quality Assurance	Celeste Annovazzi	Gente	Suprati

CPC Biotech s.r.l.	Tel. +39 039 6898029		
Via per Ossona, 22	Fax. +39 039 6899774	PMS_79001209	Page 1/1
20010 Casorezzo (MI) Italy	www.cpcbiotech.it		_