

## **CERTIFICATE OF ANALYSIS Nr 52**



**Product:** 

BioCapt® Single-Use Manufactured by Neomed srl

Product Code: 79001201

## Media: TRYPTONE SOYA AGAR

Lot Nr:	Date of Manufacturing	Expiry date :	Storage Conditi	ons:
17A2503	25 Jan 2017	25 Oct 2017	2-25°C up to exp	ire date
Sterilization: ≥25 KGy gamma irradiated		Irradiation Certificate Nr: 1217		Conforms

## CHEMICAL / PHYSICAL TESTS:

Appearance:	light amber coloured, clear to slightly opalescent gel forms in BioCapt device plates		Conforms
pH specification:	7.3 ± 0.2 at 25°C	pH result: 7,3	Conforms
Filling media Volume: 30 ml ± 1ml			Conforms
Media Composition: TSA- BSE free (see tecnichal data sheet cod 79001201)			Conforms

## MICROBIOLOGICAL TEST CONTROLS:

Sterility Control - Absence of co	ntamination	
At the following incubation condition	is (performed before gamma irradiation):	Conforms
32.5 ± 2.5	°C aerobic for 5 days	conforms
22.5 ± 2.5	°C aerobic for 5 days	

Fertility - Growth Promotion test (performed after gamma irradiation):				Conforms	
Control strain	Medium inoculation level	Incubation Conditions	Recovery Specifications	Recovery Results	Cultural Response
<i>E. coli</i> ATCC 8739	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	98,8%	Conforms
<i>P. aeruginosa</i> ATCC 9027	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	97,4%	Conforms
<i>S. aureus</i> ATCC 6538	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	93,5%	Conforms
<i>B. subtilis</i> ATCC 6633	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	95,2%	Conforms
<i>C. albicans</i> ATCC 10231	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	93,8%	Conforms
<i>A. brasiliensis</i> ATCC 16404	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	87,0%	Conforms

**GROWTH PROMOTION TEST** IN ACCORDANCE TO HARMONISED EP – USP PHARMACOPOEIAS CHAPTER, BY QUANTITATIVE SURFACE PLATING METHOD. **RECOVERY**: RESULTS AS CFU ENUMERATION ON TEST BATCH AGAR PLATES X 100/CFU ENUMERATION ON REFERENCE TESTS.

**REFERENCE MEDIUM** FOR RECOVERY TESTS: TRYPTIC SOYA AGAR FOR ALL BACTERIA AND SABOURAUD DEXTROSE AGAR FOR YEAST & MOLD SPECIES. **ATCC** IS REGISTERED TRADE MARK OF AMERICAN TYPE CULTURE COLLECTION.

Lot Status: Approved	Lot Approval Date: 6	Mar 2017
Quality Control Manager: Dott.ssa Cinzia Bovio	Signature:	Cipliatorio

