

Product	Batch	Manufacturing date	Expiration date
BioCapt Tryptic Soy Agar + Neutralizers + Pyruvate	071318034	2018/07/13	2019/07/13
Ref. 19103S (79001206)			

Gamma Rays Treatment	Irradiation Dose (kGy)	Date
Treatment Certificate No. 211.903	12,5	2018.08.27

Physical quality control	Specification	Results
Expected pH-value (25°C)	7.3 ± 0.2	7.3
Appearance of medium	Slightly opalescent	Conforms
Colour of medium	Light amber	Conforms
Filling media volume	30 ml ± 1 ml	Conforms
Media Composition	BSE free	Conforms

Sterility Control- Absence of contamination (before gamma irradiation)		
32.5 ± 2.5 °C aerobic for 7 days		Conforms
22.5 ± 2.5 °C aerobic for 7 days		Conforms

Microbiological Performance- Growth Promotion Test (after gamma irradiation)
Tested according to CLSI M22-A3, EN ISO 11133, Harmonized USP/EP/JP
Reference media: Tryptic Soy Agar

Productivity, Method of control: Quantitative				
Inoculum: 10-100 CFU				
Control strains	Incubation	Expected Results	Recovery	Results
<i>Bacillus subtilis</i> ATCC® 6633	Aerobic, 24-72 h 32.5 ± 2.5°C	70% ≤ P _R % ≤ 200%	82%	Conforms
<i>Escherichia coli</i> ATCC® 8739	Aerobic, 24-72 h 32.5 ± 2.5°C	70% ≤ P _R % ≤ 200%	124%	Conforms
<i>Staphylococcus aureus</i> ATCC® 6538	Aerobic, 24-72 h 32.5 ± 2.5°C	70% ≤ P _R % ≤ 200%	153%	Conforms
<i>Pseudomonas aeruginosa</i> ATCC® 9027	Aerobic, 24-72 h 32.5 ± 2.5°C	70% ≤ P _R % ≤ 200%	107%	Conforms
<i>Candida albicans</i> ATCC® 10231	Aerobic, 72- 120 h 32.5 ± 2.5°C	70% ≤ P _R % ≤ 200%	90%	Conforms
<i>Aspergillus brasiliensis</i> ATCC® 16404	Aerobic, 72- 120 h 32.5 ± 2.5°C	70% ≤ P _R % ≤ 200%	100%	Conforms

RECOVERY: RESULTS AS CFU ENUMERATION ON TEST BATCH AGAR PLATES x 100/ CFU ENUMERATION ON REFERENCE TESTS.
REFERENCE MEDIUM FOR RECOVERY TESTS: TSA FOR ALL BACTERIA AND SDA FOR YEAST AND MOLD SPECIES
 ATCC IS REGISTERED TRADE MARK OF AMERICAN TYPE CULTURE COLLECTION

Batch Release

Approved

Date

2018.09.03

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.



ATTESTATO DI TRATTAMENTO *Treatment Declaration*

N° 211,903
No

Data 27/08/2018
Dated

N° Rif. Lotto di Trattamento 7221/2018
Rif Batch No.

W.O. 2271241

S.O. 2035911

Mandante LIOFILCHEM S.R.L. (6)
Sender

Documento di trasporto N.° 11506
Delivery note No.

Del 01/08/2018
Dated

N° colli 8
No of Batches

Ricevente LIOFILCHEM S.R.L. (6)
Consignee

Contratto Commerciale N° 2346/2015
Contract No.

Cliente LIOFILCHEM S.R.L. (6)
Customer

Codice Packing-List www_6_61470
Packing-List code

Descrizione Prodotto (Come indicato nel form della Packing-List elettronica inviata dal Cliente) *Description of goods (As declared in the digital Packing-List sent by the Customer)*

071318034 - 071118079 - 071718111 - 072518105 - 071318082 - 062518027 - 053018043 - 072018068 - 071718002
071118080 - 062518027 - 053018043 - 071318098

Si dichiara che la merce sopra descritta è stata sottoposta a trattamento mediante esposizione a raggi gamma generati da Cobalto-60 utilizzando impianti e procedure di controllo conformi alle norme UNI EN ISO 9001, UNI CEI EN ISO 13485, UNI EN ISO 11137 e GMP.

We hereby declare that the above-mentioned goods have been irradiated with Cobalt-60 gamma rays using machinery and control procedures validated in accordance with UNI EN ISO 9001, UNI CEI EN ISO 13485, UNI EN ISO 11137 e GMP.

Sterigenics Italy si esime dalla responsabilità di eventuali errori di quanto indicato dal Cliente nella Packing-List elettronica
Sterigenics Italy will not be held responsible for any mistake in what indicated by the Customer in the digital Packing-List

Il processo di trattamento è stato effettuato secondo le specifiche del Capitolato di Fornitura: N°:

681/14 del 29/07/14
No. dated

Ciclo standard: 12.5 kGy
Standard Cycle

Dosimetro tipo Rad Perspex 4034 Lotto: PJ
Dosimeter Batch

Sterigenics Italy S.p.A - Via Marzabotto, 4 - 40061 MINERBIO (Bologna - Italia)
Tel. +39 051 66 05 998 - Fax +39 051 66 05 574 - www.sterigenics.com

Soc. Uninominale - Capitale Sociale € 2.500.000,00 I.v. - P.Iva
IT00500911201 C.F. e Reg. Imprese di Bologna 00307110379 - R.E.A.
Bologna 194057 Soggetta a Direzione e Coordinamento di Deross