



Product	Batch	Manufacturing date	Expiration date
<b>BioCapt Tryptic Soy Agar + Neutralizers + Lactamator 1000IU</b>	121817076	2017/12/18	2018/09/14

Ref. 19104S (79001210)

Gamma Rays Treatment	Irradiation Dose (kGy)	Date
Treatment Certificate No. 205.487	12,5	2018/01/25

Physical quality control	Specification	Results
Expected pH-value (25°C)	7.3 ± 0.2	7.2
Appearance of medium	Slightly opalescent	Conforms
Colour of medium	Light amber	Conforms
Filling media volume	30 ml ± 1 ml	Conforms
Media Composition	BSE free	Conforms

**Sterility Control- Absence of contamination** (before gamma irradiation)

32.5 ± 2.5 °C aerobic for 7 days	Conforms
22.5 ± 2.5 °C aerobic for 7 days	Conforms

**Microbiological Performance- Growth Promotion Test** (after gamma irradiation)

Tested according to CLSI M22-A3, EN ISO 11133, Harmonized USP/EP/JP  
Reference media: Tryptic Soy Agar

**Productivity, Method of control: Quantitative**  
Inoculum: 10-100 CFU

Control strains	Incubation	Expected Results	Recovery	Results
<i>Bacillus subtilis</i> ATCC® 6633	Aerobic, 24-72 h 32.5 ± 2.5°C	70% ≤ P <sub>R</sub> % ≤ 200%	105%	Conforms
<i>Escherichia coli</i> ATCC® 8739	Aerobic, 24-72 h 32.5 ± 2.5°C	70% ≤ P <sub>R</sub> % ≤ 200%	81%	Conforms
<i>Staphylococcus aureus</i> ATCC® 6538	Aerobic, 24-72 h 32.5 ± 2.5°C	70% ≤ P <sub>R</sub> % ≤ 200%	84%	Conforms
<i>Pseudomonas aeruginosa</i> ATCC® 9027	Aerobic, 24-72 h 32.5 ± 2.5°C	70% ≤ P <sub>R</sub> % ≤ 200%	98%	Conforms
<i>Candida albicans</i> ATCC® 10231	Aerobic, 72- 120 h 32.5 ± 2.5°C	70% ≤ P <sub>R</sub> % ≤ 200%	92%	Conforms
<i>Aspergillus brasiliensis</i> ATCC® 16404	Aerobic, 72- 120 h 32.5 ± 2.5°C	70% ≤ P <sub>R</sub> % ≤ 200%	86%	Conforms

RECOVERY: RESULTS AS CFU ENUMERATION ON TEST BATCH AGAR PLATES x 100/ CFU ENUMERATION ON REFERENCE TESTS.

REFERENCE MEDIUM FOR RECOVERY TESTS: TSA FOR ALL BACTERIA AND SDA FOR YEAST AND MOLD SPECIES

ATCC IS REGISTERED TRADE MARK OF AMERICAN TYPE CULTURE COLLECTION

**Enzymatic Activity, Method of control: Disc diffusion** Inoculum: 10<sup>4</sup>-10<sup>5</sup> CFU

Control strains	Incubation	Antimicrobial agent	Specification	Results
<i>Staphylococcus aureus</i> ATCC® 6538	Aerobic, 24 h, 32.5 ± 2.5°C	Penicillin G, P 10 IU	No inhibition	Conforms
		Ampicillin, AMP 10 µg	No inhibition	Conforms
		Cefotaxime, CTX 30 µg	No inhibition	Conforms
		Cefepime, FEP 30 µg	No inhibition	Conforms
		Meropenem, MRP 10 µg	No inhibition	Conforms

**Batch Release**

Approved

Date

2018/01/29

Signature

Quality Control  
(D. Vitagliano)

*Dario Vitagliano*

The results reported were obtained at the time of release.

# ATTESTATO DI TRATTAMENTO

## Treatment Declaration

N° 205.487  
No

Data 25/01/2018  
Dated

N° Rif. Lotto di Trattamento 414/2018  
Rif Batch No.

W.O. 2128602  
S.O. 1866690

Mandante LIOFILCHEM S.R.L. (6)  
Sender

Documento di trasporto N.° 871  
Delivery note No.

Del 17/01/2018  
Dated

N° colli 3  
No of Batches

Ricevente LIOFILCHEM S.R.L. (6)  
Consignee

Contratto Commerciale N° 2346/2015  
Contract No.

Cliente LIOFILCHEM S.R.L. (6)  
Customer

Codice Packing-List www\_6\_57424  
Packing-List code

**Descrizione Prodotto (Come indicato nel form della Packing-List elettronica inviata dal Cliente)**  
*Description of goods (As declared in the digital Packing-List sent by the Customer)*

121817076-121817112-121817077

Si dichiara che la merce sopra descritta è stata sottoposta a trattamento mediante esposizione a raggi gamma generati da Cobalto-60 utilizzando impianti e procedure di controllo conformi alle norme UNI EN ISO 9001, UNI CEI EN ISO 13485, UNI EN ISO 11137 e GMP.  
*We hereby declare that the above-mentioned goods have been irradiated with Cobalt-60 gamma rays using machinery and control procedures validated in accordance with UNI EN ISO 9001, UNI CEI EN ISO 13485, UNI EN ISO 11137 and GMP.*

Sterigenics Italy si esime dalla responsabilità di eventuali errori di quanto indicato dal Cliente nella Packing-List elettronica  
*Sterigenics Italy will not be held responsible for any mistake in what indicated by the Customer in the digital Packing-List*

**Il processo di trattamento è stato effettuato secondo le specifiche del Capitolato di Fornitura: N°: 681/14 del 29/07**  
*Treatment process has made in compliance with the specifications of the Supply Agreement: No. dated*

**Ciclo standard: 12,5 kGy**  
*Standard Cycle*

**Dosimetro tipo Red Perspex 4034 Lotto: NW**  
*Dosimeter Batch*

Sterigenics Italy S.p.A. - Via Marzabotto, 4 - 40061 MINERBIO (Bologna - Italia)  
Tel. +39 051 66 05 998 - Fax +39 051 66 05 574 - www.sterigenics.com

Soc. Uninominale - Capitale Sociale € 2.500.000,00 i.v. - P.Iva IT00500911201  
C.F. e Reg. Imprese di Bologna 00307110379 - R.E.A. Bologna 194057  
Soggetta a Direzione e Coordinamento di Deross Holding B.V.

