

Lot status: **APPROVED**

MODEL: BioCapt single use manufactured by CPC Biotech srl	PRODUCT CODE: 79001201
MEDIA: Tryptone Soya Agar	STORAGE CONDITIONS: at 2-25°C until expiry date
BATCH: 22B2203	EXPIRY DATE: 22 Feb 2023
DATE OF MANUFACTURING: 22 Feb 2022	

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	light amber coloured, clear to slightly opalescent gel forms in BioCapt device plates	Conforms
pH specification	7.3±0.2	Conforms
Filling media volume	30ml±1ml	Conforms
Media composition	TSA (Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer)	Conforms
Packaging	Triple Wrapped Sterile Irradiated (T.W.S.I.)	Conforms
Dose of irradiation	≥10 KGy	Conforms Irradiation certificate Nr: 2194/2022

MICROBIOLOGICAL TEST CONTROLS

- **Sterility control** (performed before gamma irradiation)

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 168 hours 22.5±2.5°C aerobic for 168 hours	No growth	Conforms

- **Fertility – Growth Promotion Test** (performed after gamma irradiation)

Control strain	Medium inoculation level	Incubation Conditions	Recovery Specifications	Recovery Results	Cultural Response
<i>E. coli</i> ATCC 8739	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	84,7%	Conforms
<i>P. aeruginosa</i> ATCC 9027	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	72,0%	Conforms
<i>S. aureus</i> ATCC 6538	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	110,9%	Conforms
<i>B. subtilis</i> ATCC 6633	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	84,3%	Conforms
<i>C. albicans</i> ATCC 10231	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	101,1%	Conforms
<i>A. brasiliensis</i> ATCC 16404	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	84,1%	Conforms
Environmental strain #1	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	114,3%	Conforms
Environmental strain #2	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	93,8%	Conforms

GROWTH PROMOTION TEST IN ACCORDANCE TO HARMONISED EP – USP PHARMACOPEIAS CHAPTER, BY QUANTITATIVE SURFACE PLATING METHOD.

RECOVERY: RESULTS AS CFU ENUMERATION ON TEST BATCH AGAR PLATES X 100/CFU ENUMERATION ON REFERENCE TESTS.

REFERENCE MEDIUM FOR RECOVERY TESTS: TRYPTONE SOYA AGAR FOR ALL BACTERIA AND SABOURAUD DEXTROSE AGAR FOR YEAST & MOLD SPECIES.

ATCC IS REGISTERED TRADE MARK OF AMERICAN TYPE CULTURE COLLECTION.

Date analysis start	22 Feb 2022	Date analysis end	13 Apr 2022
Quality Control	Giulia Brusati	<i>Giulia Brusati</i>	
Quality Assurance	Celeste Annovazzi	<i>Celeste Annovazzi</i>	